

# Clinical Validation Report

**Product Name:** COVID-19 antigen rapid test kit (Saliva)

**Manufacturer:** BEIJING BEIER BIOENGINEERING CO.,LTD

**Testing agency:** Tianjin Haihe Hospital



# CONTENT

<b>Clinical Validation Report.....</b>	<b>1</b>
<b>CONTENT.....</b>	<b>2</b>
<b>Introduction.....</b>	<b>3</b>
<b>1. The protocol for clinical experiment.....</b>	<b>3</b>
1.1 Basic information for clinical experiment.....	3
1.2 The design of the clinical experiment.....	4
1.2.2 The information of the kit to be evaluated.....	4
1.2.3 Information of the comparator method.....	4
1.3 Requirement for the sample collection.....	5
1.4 Inclusion/exclusion criteria of sample.....	5
1.5 The requirements about the sample number.....	5
1.6 Sample confirmation method:.....	6
1.7 The rule of the sample randomization and the blind method.....	6
1.8 Statistical method.....	6
1.9 Acceptable criteria.....	7
1.10The operation method.....	7
1.11 The train for clinical experiment.....	9
1.12 Quality control method the clinical experiment.....	9
1.13 Notes on ethical matters.....	9
<b>2.Clinical Performance study report.....</b>	<b>10</b>
2.1 The implemented time of the clinical experiment.....	10
2.2 The information of the reagent and the instrument.....	10
2.3 The distribution of the enrolled samples.....	10
2.4 The Statistical analysis results.....	11
2.5 Conclusion.....	12
<b>Appendix A-Information For samples of Clinical Trace.....</b>	<b>13</b>

## **Introduction**

A novel coronavirus (2019-nCoV) was identified in December 2019, which has resulted in hundreds of thousands of confirmed human infections worldwide. On February 11, 2020 the International Committee for Taxonomy of Viruses (ICTV) renamed the virus SARS-CoV-2. The symptoms of COVID-19 are similar to other viral respiratory diseases and include fever, cough, shortness of breath. The SARS-CoV-2 virus has four structural proteins known as S (spike), E (envelope), M (membrane) and N (nucleocapsid) proteins; the N protein holds the RNA genome and the S, E and M proteins together form the viral envelope.

Currently, the research on the Company's Covid-19 Antigen Rapid Test Kit (Saliva ) has been completed. In order to validate its clinical suitability and accuracy, we are prepared to carry out clinical validation. Entrusted by Beijing Beier Bioengineering Co.,ltd,Tianjin Haihe Hospital undertook the clinical trial on the Covid-19 Antigen Rapid Test Kit (Saliva) produced hereby in the clinical study.

## **1. The protocol for clinical experiment**

### **1.1 Basic information for clinical experiment**

#### **1.1.1 The purpose for clinical experiment**

To evaluate the diagnosis Performance of COVID-19 Antigen Rapid Test kit(Saliva)manufactured by BEIJING BEIER BIOENGINEERING CO., LTD, take the kit of Novel Coronavirus(2019-nCoV) Real Time RT-PCR Kit (Real-time Fluorescent PCR method) Manufactured by Shanghai ZJ Bio-Tech Co., Ltd. as comparator method. Test the samples with the kit manufactured by BEIJING BEIER BIOENGINEERING CO., LTD and the comparator method in parallel, and then evaluate the Percent Positive Agreement and Negative Percent Agreement of the kit to be evaluated.

#### **1.1.2 The requirement for Principal Investigator**

The Principal Investigator shall be profession clinical Inspector, and trained prior to clinical implementation.

They are at least two operators in each clinical experiment sit, one operator to number

the sample, another to test the sample, and record the result.

### **1.1.3 The planned time for clinical experiment**

2020.10~ 2021.1

## **1.2 The design of the clinical experiment**

### **1.2.1 Clinical experiment site**

The clinical trail is conducted by Tianjin Haihe Hospital. As the applicant, Beijing Beier Bioengineering Co.,Ltd was responsible for communication and contact during the clinical trail.

### **1.2.2 The information of the kit to be evaluated**

Product name: COVID-19 Antigen Rapid Test kit(Saliva)

Manufacturer: BEIJING BEIER BIOENGINEERING CO., LTD

### **1.2.3 Information of the comparator method**

Product Name: Novel Coronavirus (2019-nCoV) Real Time RT-PCR Kit (Real-time Fluorescent PCR method).

Approved by NMPA (Registration Number: GXZZ 20203400057)

Approved CE marked

Manufacturer: Shanghai ZJ Bio-Tech Co., Ltd.

The Limit of Detection (LoD) of the comparator is: 1x10<sup>3</sup> copies / mL The analysis specificity of the comparator is:

The comparator has no cross-reaction with influenza A (H1N1), influenza B (Yamagata), Respiratory syncytial virus (type A), adenovirus (type 2), Parainfluenza virus (type 1), Mycoplasma pneumoniae, Chlamydia pneumoniae, Bordetella pertussis, Streptococcus pneumoniae, rhinovirus (type A), Legionella pneumophila, Middle East Respiratory Syndrome coronavirus (MERSr COV), human coronavirus HCoV-299E, human coronavirus HCoV-HKU1, human coronavirus HCoV-NL63, human coronavirus HCoV-OC43, influenza A virus (H3N2), influenza A virus (H3N2), influenza A virus (H5N1), influenza A virus (H7N9), influenza B virus (Victoria), Human Metapneumovirus, EB virus, measles virus, human cytomegalovirus, rotavirus, norovirus, Staphylococcus aureus, Streptococcus pneumoniae, Klebsiella pneumoniae, Streptococcus pyogenes.

### **1.2.4 The instrument required for the comparator method**

Product Name: MIC qPCR Cycler

Approved by NMPA (Registration Number GXZZ 20203220013)

Manufacturer: Shanghai ZJ Bio-Tech Co., Ltd.

### **1.3 Requirement for the sample collection**

#### 1) Method for the sample collection Saliva Swab Specimen Collection

1. Place the swab against cheek, inside of mouth. Using moderate pressure, rotate the swab 10-15 times on the inside of each cheek.
2. Next, pool saliva into the mouth and allow the swab to rest at the bottom of the mouth, under the tongue, for 1-2 minutes until it is saturated with saliva.

#### **Specimen Transport and Storage:**

Freshly collected specimen should be processed as soon as possible. Collected swabs can be stored at -70°C for long time.

### **1.4 Inclusion/exclusion criteria of sample**

#### 1) Inclusion criteria

- a) specimens from patients suspected infection of SARS-CoV-2 or other respiratory viruses.
- b) Specimens positive in SARS-CoV-2 confirmed by nucleic acid assay
- c) Specimens negative in SARS-CoV-2 confirmed by nucleic acid assay from different symptoms.
- d) Specimens from the patient with the symptoms of fever, cough, shortness of breath, and dyspnea.
- e) the information of specimens is integrated.
- f) the specimens are not limited in sex, age, territory, race.
- g) the specimen types are saliva swabs.

#### 2) Exclusion criteria:

- a.) specimen contaminated by microorganism;
- b.) the specimen's storage condition not conform to the requirement state in section 4.4.
- c.) the specimen volume is insufficient for the test.
- d.) the information of the specimen is not integrated
- e.) the specimen type is not what state in section 4.4.
- f.) any specimen that the Principal Investigator think it not be selected into the experiment.

#### 3) Rejection criteria

- a.) the information of specimen is incorrect.
- b.) there is an error in the sample test, but the residual sample volume is not enough to conduct the test again.

### **1.5 The requirements about the sample number**

The samples number positive in SARS-CoV-2 should not less than 100.

The samples negative in SARS-CoV-2 should not less than 200.

## 1.6 Sample confirmation method:

The samples infection status is confirmed by clinical symptoms (fever and / or respiratory symptoms) and RT-PCR test results.

## 1.7 The rule of the sample randomization and the blind method

The experiment is planned to conduct in method of single blind experiment. the samples are numbered randomly by hospital, the rules for clinical sample numbering were set by the clinical trials site.

The sample numbered randomly by one operator, test and record by another operator.

## 1.8 Statistical method

The results of the kit to be evaluated and the comparator method were expressed by table listed below (table 1). The Percent Positive Agreement, Negative Percent Agreement, total Percent Agreement, confidence interval and kappa value (K) for the kit to be evaluated to the comparator method were calculated.

Table 1 Test results of kit to be evaluated and comparator method

kit to be evaluated	comparator method		Total
	Positive (+)	Negative (-)	
Positive (+)	a	b	a+b
Negative (-)	c	d	c+d
Total	a+c	b+d	a+b+c+d

1)Positive Percent Agreement:  $a / (a + c) \times 100 \%$

The 95% confidence interval of the Positive Percent Agreement:  $Q1=2 \times a + 1.96^2$ ,

$$Q2 = 1.96 \times \sqrt{1.96^2 + \frac{4 \times a \times c}{(a + c)}}$$

$Q3=2 \times (a+c) + 1.96^2$

The lower limit of the 95% confidence interval of the Positive Percent Agreement:

$(Q1-Q2) / Q3 \times 100\%$

The upper limit of the 95% confidence interval of the Positive Percent Agreement:

$(Q1+Q2)/Q3 \times 100\%$

2)Negative Percent Agreement:  $d / (b + d) \times 100 \%$

The 95% confidence interval of the Negative Percent Agreement:

$Q1=2 \times d + 1.96^2$ ,

$$Q2 = 1.96 \times \sqrt{1.96^2 + \frac{4 \times d \times b}{(d + b)}}$$

$$Q3 = 2 \times (d + b + 1.96^2)$$

The lower limit of the 95% confidence interval of the Negative Percent

Agreement:  $(Q1 - Q2) / Q3 \times 100\%$

The upper limit of the 95% confidence interval of the Negative Percent Agreement:

$$(Q1 + Q2) / Q3 \times 100\%$$

3) Total Percent Agreement:  $(a + d) / (a + b + c + d)$

$\times 100\%$  The 95% confidence interval of the total Percent Agreement:

$$Q1 = 2 \times (a + d) + 1.96^2,$$

$$Q2 = 1.96 \times \sqrt{1.96^2 + \frac{4 \times (a + d) \times (b + c)}{n}}$$

$$Q3 = 2 \times (n + 1.96^2)$$

The lower limit of the 95% confidence interval of the total Percent Agreement:

$$(Q1 - Q2) / Q3 \times 100\%$$

The upper limit of the 95% confidence interval of the total Percent Agreement:

$$(Q1 + Q2) / Q3 \times 100\%$$

4) kappa value (K) is calculated with the formula:

$$\text{Kappa(K)} = \frac{2(ad - bc)}{(a + b)(b + d) + (a + c)(c + d)}$$

## 1.9 Acceptable criteria

Positive Percent Agreement should  $\geq 80\%$ ;

Negative Percent Agreement should  $\geq 98\%$ ;

The total Percent Agreement should  $\geq 90\%$ ; Kappa should  $\geq 0.9$ ;

## 1.10 The operation method

### 1.10.1 The operation method of the kit to be evaluated

Allow the Test cassette, Sample lysis and specimens to equilibrate to temperature (15-30°C or 59-86°F) before testing.

- 1) Bring the Test cassette to room temperature before opening the foil pouch. The test cassette should be used within 1 hour after take out from the sealed pouch.
- 2) Place the test cassette on a clean and level surface.
- 3) Take the test according to instruction strictly.
- 4) Wait for the colored line(s) to appear. Read the results in 15 -20 minutes. Do not interpret the result exceed 20 minutes.

### 1.10.2 The operation method of the comparator method

#### 1) RNA-Extraction

Take Different brand RNA extraction kits are available. You may use your own extraction systems or the commercial kits based on the yield. For the RNA extraction, please follow the manufacturer's instructions. The recommended extraction kits are as follows:

Nucleic Acid Isolation Kit	Cat. Number	Manufacturer
RNA Isolation Kit	ME-0010/ME-00I2 1	ZJ Biotech

It is noted that the negative control in this kit should be nucleic acid extracted with the same protocol for specimens. The positive control doesn't need to be nucleic acid extracted.

#### 2) internal Control

The internal control in this kit should be added into the extraction mixture with 1µl/rxn to monitor the whole process.

#### 3) RT-PCR Protocol

The Master Mix volume for each reaction should be pipetted as follows:

The volumes of Super Mix and Enzyme Mix per reaction multiply with the number of samples, which includes the number of controls, standards, and samples prepared.

Molecular Grade Water is used as the negative control. For reasons of imprecise pipetting, always add an extra virtual sample. Mix completely and then spin down briefly with a centrifuge.

Pipet 20µl Master Mix with micropipettes of sterile filter tips to each of the Real Time PCR reaction well. add 5µl sample (nucleic acid extracted from negative control and specimen, positive control with no extraction) to different well respectively. close the plates/tubes with the cap immediately. to avoid contamination.

Spin down briefly to collect the Master Mix in the bottom of the reaction tubes.

#### 4) Perform the following protocol in the instrument MIC POC Dx48:

Selection of Fluorescence Channels	
FAM	ORF1ab
HEX/VIC/JOE	Gene N
Cal Red 610/ROX/TEXAS RED	GeneE



Cy5	IC
-----	----

45°C for 10min	1 cycle
95°C for 90sec	1 cycle
95°C for 3sec, 58°C for 20sec (Fluorescence measured at 58°C)	45cycles

5)Threshold Setting: Just above the maximum level of molecular grade water.

6)Quality Control: Negative Control and Positive Control must be performed correctly; otherwise the sample results are invalid.

Control Channel	Ct value			
	FAM(Gene ORF1ab)	HEX/VIC/JOE (Gene N)	Cal Red 610 (Gene E)	Cy5
Negative control	UNDET	UNDET	UNDET	25-40
Positive control	≤35	≤35	≤35	≤35

### 1.11 The train for clinical experiment

To ensure that the clinical experiment can be implemented correctly, the operation train was provided by BEIJING BEIER BIOENGINEERING CO., LTD prior to the clinical trial.

### 1.12 Quality control method the clinical experiment

The CRC will inspect of clinical trials periodically. The main work of the CRC includes:

- 1)whether the clinical trial is carried out according to the clinical experiment protocol;
- 2)whether record of the date is consistent with the test result;
- 3)whether the record of the date is integrated

### 1.13 Notes on ethical matters

The specimens involved in this clinical trial are all the specimens for the clinical institution's test, has nothing risk to the subjects, and the test results of this trial will not be reported to the subjects. Therefore, the subjects will not face any risk of false negative or false positive results. The subjects also did not need to do additional examination, and did not need to pay additional fees. In the clinical trial, we only analyze the data of the enrolled specimens, we will check the original data only to

confirm the authenticity of the specimen source, code all the enrolled specimens, hide the name, medical record number and other information of the subject, and fully guarantee the privacy of the subject. so we apply to the Ethics Committee for exemption from informed consent.

## **2.Clinical Performance study report**

### **2.1 The implemented time of the clinical experiment**

2020.10.6~ 2021.1.22

### **2.2 The information of the reagent and the instrument**

#### **2.2.1 The information of the kit to be evaluated**

Product name: COVID-19 antigen rapid test kit(Saliva)  
Manufacturer: BEIJING BEIER BIOENGINEERING CO.,LTD  
Model: 20Tests / Kit  
Lot:20200801  
Date of expiry: 04/30/2022

#### **2.2.2 Information of the comparator method**

Product Name: Novel Coronavirus(2019-nCoV) Real Time RT-PCR Kit (Real-time Fluorescent PCR method).  
Manufacturer: Shanghai ZJ Bio-Tech Co., Ltd. Registration No: GXZZ20203400057  
Model: 25Tests / kit  
Lot: P20200801  
Date of expiry: 01/31/2021

The reagent required but not provide for the comparator method Product Name:  
Autrax automatic nucleic acid detection system  
Registration No : HXZZ 20172410039  
manufacturer : Shanghai ZJ Bio-Tech Co., Ltd.  
Model: Autrax192

The instrument required for the comparator method Product Name: MIC qPCR Cycler  
Approved by NMPA (Registration No:GXZZ 20203220013 )  
Manufacturer: Shanghai ZJ Bio-Tech Co., Ltd

### **2.3 The distribution of the enrolled samples**

There are 303 negative clinical samples and 171 positive clinical samples collected from patients with signs and symptoms of an upper respiratory infection. Each of clinical swab samples by Saliva swab, and confirmed by RT-PCR The information

about sample is listed in table below.

Table : State of sample

State of sample	Saliva swab
Positive	171
Negative	303

## 2.4 The Statistical analysis results

### 2.4.1 The results of evaluated antigen kit and comparator nucleic acid kit of positive sample

Evaluated antigen kit		Comparator nucleic acid kit	
Positive	negative	CT value	number
76	0	<25	76
65	0	25 <CT < 28	65
65	0	25 <CT < 28	65
23	3	28 <CT < 31	26
1	3	31 <CT < 34	4
165	6		171

### 2.4.2 The results of evaluated antigen kit and comparator nucleic acid kit of negative sample

evaluated antigen kit		comparator nucleic acid kit	
Positive	negative	CT value	Number
1	302	Negative	303

### 2.4.3 Coincidence rate between evaluated antigen kit and comparator nucleic acid kit

Reagents	Clinical diagnosis (nucleic acid Reagents)		Total
	Positive	Negative	

Beier Reagents	Positive	165	1	166
	Negative	6	302	308
Total		171	303	474

Results analysis:

Sensitivity=96.5%(95% CI: 93.7%~99.3%)

Specificity=99.7%(95% CI: 99.0% -100%)

Positive predictive values= 99.4%(95% CI: 98.2% -100%)

Negative predictive values= 98.0%(95% CI: 96.5%-99.6%)

Total consistent: 98.5% (95%CI: 97.4%~99.6%)

## 2.5 Conclusion

Compared with the Test Kit and Novel Coronavirus(2019-nCoV) Real Time RT-PCR Kit (Real-time Fluorescent PCR method) Manufactured by Shanghai ZJ Bio-Tech Co,Ltd. The test results show that the diagnostic specificity, diagnostic sensitivity, total coincidence rate percentage of the Test Kit are relatively high, meet the requirement of the test kit.

## Appendix A-Information For samples of Clinical Trace

Serial number	Sample Type Antigen	Beier Kit Result	COVID-19 Diagnostic	Name of RT-PCR	RT-PCR Result
1	Saliva	+	+	MIC-qPCR	+
2	Saliva	+	+	MIC-qPCR	+
3	Saliva	+	+	MIC-qPCR	+
4	Saliva	+	+	MIC-qPCR	+
5	Saliva	+	+	MIC-qPCR	+
6	Saliva	+	+	MIC-qPCR	+
7	Saliva	+	+	MIC-qPCR	+
8	Saliva	+	+	MIC-qPCR	+
9	Saliva	+	+	MIC-qPCR	+
10	Saliva	+	+	MIC-qPCR	+
11	Saliva	+	+	MIC-qPCR	+
12	Saliva	+	+	MIC-qPCR	+
13	Saliva	+	+	MIC-qPCR	+
14	Saliva	+	+	MIC-qPCR	+
15	Saliva	+	+	MIC-qPCR	+
16	Saliva	+	+	MIC-qPCR	+
17	Saliva	+	+	MIC-qPCR	+
18	Saliva	+	+	MIC-qPCR	+
19	Saliva	+	+	MIC-qPCR	+
20	Saliva	+	+	MIC-qPCR	+
21	Saliva	-	+	MIC-qPCR	+
22	Saliva	+	+	MIC-qPCR	+
23	Saliva	+	+	MIC-qPCR	+
24	Saliva	+	+	MIC-qPCR	+
25	Saliva	+	+	MIC-qPCR	+
26	Saliva	+	+	MIC-qPCR	+
27	Saliva	+	+	MIC-qPCR	+
28	Saliva	+	+	MIC-qPCR	+
29	Saliva	+	+	MIC-qPCR	+
30	Saliva	+	+	MIC-qPCR	+
31	Saliva	+	+	MIC-qPCR	+
32	Saliva	+	+	MIC-qPCR	+
33	Saliva	+	+	MIC-qPCR	+

34	Saliva	+	+	MIC-qPCR	+
35	Saliva	+	+	MIC-qPCR	+
36	Saliva	+	+	MIC-qPCR	+
37	Saliva	+	+	MIC-qPCR	+
38	Saliva	+	+	MIC-qPCR	+
39	Saliva	+	+	MIC-qPCR	+
40	Saliva	+	+	MIC-qPCR	+
41	Saliva	+	+	MIC-qPCR	+
42	Saliva	+	+	MIC-qPCR	+
43	Saliva	+	+	MIC-qPCR	+
44	Saliva	+	+	MIC-qPCR	+
45	Saliva	+	+	MIC-qPCR	+
46	Saliva	+	+	MIC-qPCR	+
47	Saliva	+	+	MIC-qPCR	+
48	Saliva	+	+	MIC-qPCR	+
49	Saliva	+	+	MIC-qPCR	+
50	Saliva	+	+	MIC-qPCR	+
51	Saliva	+	+	MIC-qPCR	+
52	Saliva	-	+	MIC-qPCR	+
53	Saliva	+	+	MIC-qPCR	+
54	Saliva	+	+	MIC-qPCR	+
55	Saliva	+	+	MIC-qPCR	+
56	Saliva	+	+	MIC-qPCR	+
57	Saliva	+	+	MIC-qPCR	+
58	Saliva	+	+	MIC-qPCR	+
59	Saliva	+	+	MIC-qPCR	+
60	Saliva	+	+	MIC-qPCR	+
61	Saliva	+	+	MIC-qPCR	+
62	Saliva	+	+	MIC-qPCR	+
63	Saliva	+	+	MIC-qPCR	+
64	Saliva	+	+	MIC-qPCR	+
65	Saliva	+	+	MIC-qPCR	+
66	Saliva	+	+	MIC-qPCR	+
67	Saliva	+	+	MIC-qPCR	+
68	Saliva	+	+	MIC-qPCR	+
69	Saliva	+	+	MIC-qPCR	+
70	Saliva	-	+	MIC-qPCR	+
71	Saliva	+	+	MIC-qPCR	+

72	Saliva	+	+	MIC-qPCR	+
73	Saliva	+	+	MIC-qPCR	+
74	Saliva	+	+	MIC-qPCR	+
75	Saliva	+	+	MIC-qPCR	+
76	Saliva	+	+	MIC-qPCR	+
77	Saliva	+	+	MIC-qPCR	+
78	Saliva	+	+	MIC-qPCR	+
79	Saliva	+	+	MIC-qPCR	+
80	Saliva	+	+	MIC-qPCR	+
81	Saliva	+	+	MIC-qPCR	+
82	Saliva	+	+	MIC-qPCR	+
83	Saliva	+	+	MIC-qPCR	+
84	Saliva	+	+	MIC-qPCR	+
85	Saliva	-	+	MIC-qPCR	+
86	Saliva	+	+	MIC-qPCR	+
87	Saliva	+	+	MIC-qPCR	+
88	Saliva	+	+	MIC-qPCR	+
89	Saliva	+	+	MIC-qPCR	+
90	Saliva	+	+	MIC-qPCR	+
91	Saliva	+	+	MIC-qPCR	+
92	Saliva	+	+	MIC-qPCR	+
93	Saliva	+	+	MIC-qPCR	+
94	Saliva	+	+	MIC-qPCR	+
95	Saliva	+	+	MIC-qPCR	+
96	Saliva	+	+	MIC-qPCR	+
97	Saliva	+	+	MIC-qPCR	+
98	Saliva	+	+	MIC-qPCR	+
99	Saliva	+	+	MIC-qPCR	+
100	Saliva	+	+	MIC-qPCR	+
101	Saliva	-	+	MIC-qPCR	+
102	Saliva	+	+	MIC-qPCR	+
103	Saliva	+	+	MIC-qPCR	+
104	Saliva	+	+	MIC-qPCR	+
105	Saliva	+	+	MIC-qPCR	+
106	Saliva	+	+	MIC-qPCR	+
107	Saliva	+	+	MIC-qPCR	+
108	Saliva	+	+	MIC-qPCR	+
109	Saliva	+	+	MIC-qPCR	+
110	Saliva	+	+	MIC-qPCR	+

111	Saliva	+	+	MIC-qPCR	+
112	Saliva	+	+	MIC-qPCR	+
113	Saliva	+	+	MIC-qPCR	+
114	Saliva	+	+	MIC-qPCR	+
115	Saliva	+	+	MIC-qPCR	+
116	Saliva	+	+	MIC-qPCR	+
117	Saliva	+	+	MIC-qPCR	+
118	Saliva	+	+	MIC-qPCR	+
119	Saliva	+	+	MIC-qPCR	+
120	Saliva	+	+	MIC-qPCR	+
121	Saliva	+	+	MIC-qPCR	+
122	Saliva	+	+	MIC-qPCR	+
123	Saliva	+	+	MIC-qPCR	+
124	Saliva	+	+	MIC-qPCR	+
125	Saliva	+	+	MIC-qPCR	+
126	Saliva	+	+	MIC-qPCR	+
127	Saliva	+	+	MIC-qPCR	+
128	Saliva	+	+	MIC-qPCR	+
129	Saliva	+	+	MIC-qPCR	+
130	Saliva	+	+	MIC-qPCR	+
131	Saliva	+	+	MIC-qPCR	+
132	Saliva	+	+	MIC-qPCR	+
133	Saliva	-	+	MIC-qPCR	+
134	Saliva	+	+	MIC-qPCR	+
135	Saliva	+	+	MIC-qPCR	+
136	Saliva	+	+	MIC-qPCR	+
137	Saliva	+	+	MIC-qPCR	+
138	Saliva	+	+	MIC-qPCR	+
139	Saliva	+	+	MIC-qPCR	+
140	Saliva	+	+	MIC-qPCR	+
141	Saliva	+	+	MIC-qPCR	+
142	Saliva	+	+	MIC-qPCR	+
143	Saliva	+	+	MIC-qPCR	+
144	Saliva	+	+	MIC-qPCR	+
145	Saliva	+	+	MIC-qPCR	+
146	Saliva	+	+	MIC-qPCR	+
147	Saliva	+	+	MIC-qPCR	+
148	Saliva	+	+	MIC-qPCR	+
149	Saliva	+	+	MIC-qPCR	+



150	Saliva	+	+	MIC-qPCR	+
151	Saliva	+	+	MIC-qPCR	+
152	Saliva	+	+	MIC-qPCR	+
153	Saliva	+	+	MIC-qPCR	+
154	Saliva	+	+	MIC-qPCR	+
155	Saliva	+	+	MIC-qPCR	+
156	Saliva	+	+	MIC-qPCR	+
157	Saliva	+	+	MIC-qPCR	+
158	Saliva	+	+	MIC-qPCR	+
159	Saliva	+	+	MIC-qPCR	+
160	Saliva	+	+	MIC-qPCR	+
161	Saliva	+	+	MIC-qPCR	+
162	Saliva	+	+	MIC-qPCR	+
163	Saliva	+	+	MIC-qPCR	+
164	Saliva	+	+	MIC-qPCR	+
165	Saliva	+	+	MIC-qPCR	+
166	Saliva	+	+	MIC-qPCR	+
167	Saliva	+	+	MIC-qPCR	+
168	Saliva	+	+	MIC-qPCR	+
169	Saliva	+	+	MIC-qPCR	+
170	Saliva	+	+	MIC-qPCR	+
171	Saliva	+	+	MIC-qPCR	+